



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

FARO Technologies (Shanghai) Co., Ltd.
1/F, Building No. 2, 188 Pingfu Road
Juxin Information Technology Park, Xuhui District
Shanghai 200231 China

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 25 January 2028

Certificate Number: L1147.07-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

FARO Technologies (Shanghai) Co., Ltd.

1/F, Building No. 2, 188 Pingfu Road
 Juxin Information Technology Park, Xuhui District
 Shanghai 200231 China

Adrian Lim
 Michael Bai
 +86-21-61917600

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **22 January 2026**

Certificate Number: **L1147.07-1** Certificate Expiry Date: **25 January 2028**

Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Volumetric Performance	(0 to 2.2) m	3.5 µm	ASME B89.4.22-2004 at 5.2, 5.3 and 5.4 using Kinematic Scale Bar
	Effective Diameter (3 to 25.4) mm	1.0 µm	Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Size Error (PSize)	Working Volume:		ISO 10360-12:2016 at 6.2 using 25.4 mm Test Sphere
	1.5 m	2.2 µm	
	2.0 m	2.2 µm	
	2.5 m	2.8 µm	
	3.0 m	6.4 µm	
	3.5 m	7.8 µm	
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Form Error (PForm)	Working Volume:		ISO 10360-12:2016 at 6.2 using 25.4 mm Test Sphere
	1.5 m	2.1 µm	
	2.0 m	2.1 µm	
	2.5 m	2.4 µm	
	3.0 m	3.7 µm	
	3.5 m	3.7 µm	
	4.0 m	5.3 µm	

Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Articulated Location Error (LDia)	Working Volume: 1.5 m 2.0 m 2.5 m 3.0 m 3.5 m 4.0 m	7.5 μm 7.5 μm 8.7 μm 11 μm 17 μm 18 μm	ISO 10360-12:2016 at 6.3 using 25.4 mm Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Length Measurement Error, Unidirectional (EU _{ni})	(0 to 1.05) m (0 to 1.36) m (0 to 1.8) m (0 to 2.11) m (0 to 2.42) m (0 to 2.64) m	3.1 μm 3.7 μm 4.0 μm 5.7 μm 6.4 μm 8.4 μm	ISO 10360-12:2016 at 6.4 using Kinematic Scale Bar
Articulated Arm Coordinate Measuring Machines (AACMM) with Optical Distance Sensors: Articulated Location Value	 Sphere Diameter: 50.8 mm	 4.4 μm	Based on ISO 10360-08:2013 Annex D by comparison to Test Sphere
Laser Line Probe (LLP): Diameter Z Distance/Position	 Cylinder Diameter: 25.4 mm 25.4 mm 25.4 mm (75 to 360) mm (80 to 230) mm	 1.9 μm 3.8 μm 7.0 μm 1.4 μm 2.9 μm	Internal Procedure by comparison to Reference Glass Cylinder Reference Aluminum Cylinder Reference Ceramic Cylinder Reference Distance/ Position Fixture
Faro Laser Tracker: Ranging Length Measurement	(0.04 to 25) m	(2 + 0.4L) μm	ASME B89.4.19-2006 using Reference Laser Tracker
FARO Laser Tracker: Orientation Error of Six-DOF Probe	(2.5 to 10) m	2.0 μm	ISO 10360-10:2016

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = Length in meters, X = the perpendicular distance from the tracker to the space frame.



Jason Stine, Vice President

