



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**FARO Mexico - 3D Measurement Technologies  
S. de R.L. de C.V.  
215 Avenida Centuria,  
Parque Industrial Milenium  
Apodaca, Nuevo Leon, 66600 Mexico**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 25 January 2026

Certificate Number: L1147.11-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**FARO Mexico - 3D Measurement Technologies S. de R.L. de C.V.**

215 Avenida Centuria,  
Parque Industrial Milenium  
Apodaca, Nuevo Leon, 66600 Mexico  
Rachel Sowers  
407-333-9911

**CALIBRATION**

Valid to: **January 25, 2026**

Certificate Number: **L1147.11-1**

**Length-Dimensional Metrology**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-) <sup>2</sup></b>	<b>Reference Standard, Method, and/or Equipment</b>
Articulated Arm Coordinate Measurement Machine (AACMM):  Volumetric Performance  Effective Diameter	(0 to 2.2) m    (3 to 25.4) mm	    3.5 µm    1.0 µm	ASME B89.4.22-2004 at 5.2, 5.3 and 5.4 ISO 10360-2:2001  Kinematic Scale Bar  Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Size Error (PSize)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 1.4 µm 6 Axis 2.0m- 1.6 µm 6 Axis 2.5m- 1.8 µm 6 Axis 3.0m- 2.4 µm 6 Axis 3.5m- 3.2 µm 6 Axis 4.0m- 4.0 µm 7 Axis 2.0m- 2.0 µm 7 Axis 2.5m- 2.2 µm 7 Axis 3.0m- 3.2 µm 7 Axis 3.5m- 4.0 µm 7 Axis 4.0m- 5.2 µm	ISO 10360-12:2016 6.2, 6.3, 6.4  Test Sphere

**Length-Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Form Error (PForm)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.4 μm 6 Axis 2.5m- 3.6 μm 6 Axis 3.0m- 5.2 μm 6 Axis 3.5m- 6.8 μm 6 Axis 4.0m- 7.6 μm 7 Axis 2.0m- 4.0 μm 7 Axis 2.5m- 4.4 μm 7 Axis 3.0m- 6.4 μm 7 Axis 3.5m- 7.8 μm 7 Axis 4.0m- 8.8 μm	Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Articulated Location Error (LDia)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.0 μm 6 Axis 2.5m- 3.2 μm 6 Axis 3.0m- 9.2 μm 6 Axis 3.5m- 13 μm 6 Axis 4.0m- 16 μm 7 Axis 2.0m- 4.0 μm 7 Axis 2.5m- 4.4 μm 7 Axis 3.0m- 15 μm 7 Axis 3.5m- 18 μm 7 Axis 4.0m- 22 μm	Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Length Measurement Error, Unidirectional (EUni)	(0 to 1.05) m (0 to 1.36) m (0 to 1.8) m (0 to 2.11) m (0 to 2.42) m (0 to 2.64) m	3.1 μm 3.7 μm 4.0 μm 5.7 μm 6.4 μm 8.4 μm	Kinematic Scale Bar
Articulated Arm Coordinate Measurement Machine (AACMM): Single Point Articulation Performance	N/A <sup>3</sup>	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.2 μm 6 Axis 2.5m- 3.6 μm 6 Axis 3.0m- 5.4 μm 6 Axis 3.5m- 7.2 μm 6 Axis 4.0m- 9.0 μm 7 Axis 2.0m- 3.6 μm 7 Axis 2.5m- 4.0 μm 7 Axis 3.0m- 6.4 μm 7 Axis 3.5m- 9.0 μm 7 Axis 4.0m- 11 μm	B89.4.22-2004 and ISO 10360-12:2016  Test Sphere


**Length-Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measuring Machines (AACMM) with Optical Distance Sensors:			Based on ISO 10360-08:2013 Annex D
Articulated Location Value	Sphere Diameter: 50.8 mm	4.4 μm	Test Sphere
Laser Line Probe (LLP):			Internal Procedure:
Diameter	Cylinder Diameter: 25.4 mm	3.8 μm	Reference Cylinder
Z Distance/Position	(80 to 230) mm	2.9 μm	Calibrated Distance/ Position by Laser Interferometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = Length in meters,  $X$  = the perpendicular distance from the tracker to the space frame.
3. Point measurements do not have a range.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L1147.11-1.



Jason Stine, Vice President