



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

FARO Technologies UK Ltd
Unit 1A Great Central Way
Rugby, CV21 3XH United Kingdom

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 25 January 2026

Certificate Number: L1147.14-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

FARO Technologies UK Ltd
Unit 1A Great Central Way
Rugby, CV21 3XH United Kingdom
Forhadul Islam
+44 02476 973000

CALIBRATION

Valid to: **January 25, 2026**

Certificate Number: **L1147.14-1**

Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Volumetric Performance Effective Diameter	(0 to 2.2) m (3 to 25.4) mm	 3.5 µm 1.0 µm	ASME B89.4.22-2004 at 5.2, 5.3 and 5.4 ISO 10360-2:2001 Kinematic Scale Bar Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Size Error (PSize)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 1.4 µm 6 Axis 2.0m- 1.6 µm 6 Axis 2.5m- 1.8 µm 6 Axis 3.0m- 2.4 µm 6 Axis 3.5m- 3.2 µm 6 Axis 4.0m- 4.0 µm 7 Axis 2.0m- 2.0 µm 7 Axis 2.5m- 2.2 µm 7 Axis 3.0m- 3.2 µm 7 Axis 3.5m- 4.0 µm 7 Axis 4.0m- 5.2 µm	ISO 10360-12:2016 6.2, 6.3, 6.4 Test Sphere

Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Probing Form Error (PForm)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.4 μm 6 Axis 2.5m- 3.6 μm 6 Axis 3.0m- 5.2 μm 6 Axis 3.5m- 6.8 μm 6 Axis 4.0m- 7.6 μm 7 Axis 2.0m- 4.0 μm 7 Axis 2.5m- 4.4 μm 7 Axis 3.0m- 6.4 μm 7 Axis 3.5m- 7.8 μm 7 Axis 4.0m- 8.8 μm	Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Articulated Location Error (LDia)	Sphere Diameter: 25.4 mm	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.0 μm 6 Axis 2.5m- 3.2 μm 6 Axis 3.0m- 9.2 μm 6 Axis 3.5m- 13 μm 6 Axis 4.0m- 16 μm 7 Axis 2.0m- 4.0 μm 7 Axis 2.5m- 4.4 μm 7 Axis 3.0m- 15 μm 7 Axis 3.5m- 18 μm 7 Axis 4.0m- 22 μm	Test Sphere
Articulated Arm Coordinate Measurement Machine (AACMM): Length Measurement Error, Unidirectional (EUni)	(0 to 1.05) m (0 to 1.36) m (0 to 1.8) m (0 to 2.11) m (0 to 2.42) m (0 to 2.64) m	3.1 μm 3.7 μm 4.0 μm 5.7 μm 6.4 μm 8.4 μm	Kinematic Scale Bar
Articulated Arm Coordinate Measurement Machine (AACMM): Single Point Articulation Performance	N/A ³	6 Axis 1.5m- 2.4 μm 6 Axis 2.0m- 3.2 μm 6 Axis 2.5m- 3.6 μm 6 Axis 3.0m- 5.4 μm 6 Axis 3.5m- 7.2 μm 6 Axis 4.0m- 9.0 μm 7 Axis 2.0m- 3.6 μm 7 Axis 2.5m- 4.0 μm 7 Axis 3.0m- 6.4 μm 7 Axis 3.5m- 9.0 μm 7 Axis 4.0m- 11 μm	B89.4.22-2004 and ISO 10360-12:2016 Test Sphere


Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measuring Machines (AACMM) with Optical Distance Sensors: Articulated Location Value	Sphere Diameter: 50.8 mm	4.4 μm	Based on ISO 10360-08:2013 Annex D Test Sphere
Laser Line Probe (LLP): Diameter Z Distance/Position	Cylinder Diameter: 25.4 mm (80 to 230) mm	3.8 μm 2.9 μm	Internal Procedure: Reference Cylinder Calibrated Distance/ Position by Laser Interferometer
Faro Laser Tracker: Ranging Length Measurement	(0.04 to 25) m	(2 + 0.4L) μm	ASME B89.4.19-2006: Reference Laser Tracker
Faro Laser Tracker: Transverse Length Measurement	(0.23 to 6.2) m	(8 + 1.2X) μm	ASME B89.4.19:2006 Reference Laser Tracker Kinematic Scale Bars

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = Length in meters, X = the perpendicular distance from the tracker to the space frame.
3. Point measurements do not have a range.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L1147.14-1.



Jason Stine, Vice President