



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Faro Technologies do Brasil Ltda**  
**Rua San Jose, 360 - Parque Industrial**  
**San Jose – Cotia, S. P. 06715-862**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 25 January 2022

Certificate Number: L1147.03-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

### Faro Technologies do Brasil Ltda

Rua San Jose, 360 - Parque Industrial  
San Jose – Cotia, S. P. 06715-862  
Carlos Valentim  
+55-11-3500-4600

### CALIBRATION

Valid to: **January 25, 2022**

Certificate Number: **L1147.03-1**

#### Length-Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Articulated Arm Coordinate Measurement Machine (AACMM): Volumetric Performance	(0 to 0.9) m (0 to 2.2) m	(0.35 + 0.45L) μm 3.5 μm	ASME B89.4.22-2004 at 5.2, 5.3 and 5.4 ISO 10360-2:2001  Ball Bar Kinematic Scale Bar
Effective Diameter	(3 to 25.4) mm	1 μm	Test Sphere

**Length-Dimensional Metrology**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-) <sup>2</sup></b>	<b>Reference Standard, Method, and/or Equipment</b>
Articulated Arm Coordinate Measurement Machine (AACMM):			ISO 10360-12:2016 6.2, 6.3, 6.4
Probing Size Error (PSize)	Sphere Diameter: 25.4 mm	1 μm	Test Sphere
Probing Form Error (PForm)	Sphere Diameter: 25.4 mm	0.9 μm	Test Sphere
Articulated Location Error (LDia)	Sphere Diameter: 25.4 mm	1.7 μm	Test Sphere
Length Measurement Error, Unidirectional (EUni)	(0 to 1.05) m (0 to 1.36) m (0 to 1.8) m (0 to 2.11) m (0 to 2.42) m (0 to 2.64) m	3.1 μm 3.7 μm 4.0 μm 5.7 μm 6.4 μm 8.4 μm	Kinematic Scale Bar
Articulated Arm Coordinate Measurement Machine (AACMM):			B89.4.22-2004 and ISO 10360-12:2016
Single Point Articulation Performance	N/A <sup>3</sup>	0.41 μm	Test Sphere
Articulated Arm Coordinate Measuring Machines (AACMM) with Optical Distance Sensors:			Based on ISO 10360-08:2013 Annex D
Articulated Location Value	Sphere Diameter: 50.8 mm	4.4 μm	Test Sphere

**Length-Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Laser Line Probe (LLP):  Diameter  Z Distance/Position	Cylinder Diameter: 25.4 mm  (75 to 360) mm	  1.8 μm  1.5 μm	Internal Procedure:  Reference Cylinder  Calibrated distance/ Position by Laser Interferometer
Faro Laser Tracker:  Ranging Length Measurement	(0.04 to 25) m	(2 + 0.4L) μm	ASME B89.4.19-2006:  Reference Laser Tracker
Faro Laser Tracker:  Transverse Length Measurement	(0.23 to 6.2) m	(8 + 1.2X) μm	ASME B89.4.19:2006  Reference Laser Tracker  Kinematic Scale Bars

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = Length in meters.  $X$  = the perpendicular distance from the tracker to the space frame.
3. Point measurements do not have a range.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L1147.03-1.



R. Douglas Leonard Jr., VP, PILR SBU